



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/646,881	08/21/2003	Larry W. Ishler		1029

7590 04/04/2005

Kenneth W. Wargo, Esq.  
2222 West Grandview Boulevard  
Erie, PA 16506

EXAMINER

NASSER, ROBERT L

ART UNIT	PAPER NUMBER
----------	--------------

3736

DATE MAILED: 04/04/2005

Please find below and/or attached an Office communication concerning this application or proceeding.



UNITED STATES DEPARTMENT OF COMMERCE

U.S. Patent and Trademark Office

Address : COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

ed

APPLICATION NO./ CONTROL NO.	FILING DATE	FIRST NAMED INVENTOR / PATENT IN REEXAMINATION	ATTORNEY DOCKET NO.
---------------------------------	-------------	---	---------------------

10/646,881

EXAMINER
----------

ART UNIT	PAPER
----------	-------

03252005

DATE MAILED:

Please find below and/or attached an Office communication concerning this application or proceeding.

Commissioner for Patents

see attached

ROBERT L. NASSER  
PRIMARY EXAMINER

*Robert L. Nasser*

Robert L. Nasser  
Primary Examiner  
Art Unit: 3736

Art Unit: 3736

This supplemental action is being issued to clarify that Khalil, Wong, and Galen “do not” determine the square root of the product. The examiner inadvertently left the words “do not” out of the previous action.

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Khalil et al 2003/0023151 (paragraph 0016), Wong et al (column 3, lines 34-62) and Galen et al 2004/0265941 all discuss making fasting glucose and glycolated hemoglobin measurements, but do not determine the square root of the product and use this value as a base line to be adjusted by ear temperature differences, as recited in the claims.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Robert L. Nasser whose telephone number is (571) 272-4731. The examiner can normally be reached on Mon-Fri, variable hours.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on (571) 272-4726. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Robert L. Nasser

Application/Control Number: 10/646,881


Page 3

Art Unit: 3736

Primary Examiner  
Art Unit 3736

RLN

March 25 21, 2005

A handwritten signature in cursive script, appearing to read "Robert L. Nasser".

ROBERT L. NASSER  
PRIMARY EXAMINER